

Certificate of Analysis

Spectrophotometric Chlorophyll Standards

16 September 2017

Lot CAS157-02

Original Chlorophyll Solution A:

Mixed: 2 August 2017

Spectrophotometric Data: 16 September 2017

Abs₆₆₄ = 0.88249 Abs₇₅₀ = 2.8381 E-03

Chlorophyll concentration determination (Jeffrey Method, 1997)

Chlorophyll-a = 10.03 mg/L

Chlorophyll dilution A is provided with this note. The concentration value is guaranteed through September, 2018, but degradation depends on the handling of the standard. Standards must be kept at temperatures below -20° C and in the dark until use. Spectrophotometers should be standardized in subdued light with this solution warmed to room temperature, about 22° C. The determination of chlorophyll-a in aqueous acetone solvent is a sensitive technique. Please pay special attention to the configuration of the spectrophotometer itself. A bandpass of 2.0 nm or less is necessary to properly use this standard.



John Salinas
Analytical Chemist

16 September 2017
Date

TURNER
 **DESIGNS**

Reliable Instruments for an Unreliable World
1995 N. 1st Street Telephone 408-749-0994
San Jose, CA 95112 Toll Free 877-316-8049
www.turnerdesigns.com Fax 408-749-0998